



Puławy, May 11th 2011

Answers to the questions asked by Economic Operators concerning Public Contract no INS/LWC-20/2011.

Items 3.4 and 3.5:

Are flow ranges of 2 liquids to be identical, i.e. 0-6 l/h?

If not, I kindly request to give corrected parameters.

Answer:

Flow ranges for H₂O metering pump and H₂O₂ pump are to be identical. Vast flow range availability is required.

Item 3.8.

- a. What is air purity supposed to be?
- b. What is flow range expected?

Answer:

- a) Awarding Entity assumes that Air feed system will use the bottle with compressed air as a source of air. The parameters of such gas are as following:
 - Composition: nitrogen 79,1 %, oxygen 20,9 %.
 - Pressure in the bottle max. 150 bar,

b) min. Flow range 0 - 10 NL/h

item 3.9

Is flow range of gas mass flow meter supposed to be ONLY 2 NI/h?

Answer:

Expected flow range is 2 NL/h.

item 3.11

What type of static mixer should be offered? Is it up to offeror to decide what type to offer?

Static mixer which is in compliance with all the Specification requirements should be offered.

Question :

Due to the fact that Awarding Entity when specifying the Apparatus has proven that he has the knowledge of particular elements of such apparatus and he requires the supply of apparatus to





be complete and assembled I hereby assume that he knows how those elements are to be connected.


Therefore, I request you to send us schematic or conceptual diagram of the apparatus specified.

Answer:

Awarding Entity hereby declares that he has no schematic or conceptual diagram of apparatus for supercritical water oxidation at his disposal, and the selection of the apparatus configuration in compliance with all the norms, regulations, Specification requirements and in compliance with the intended use of the apparatus is to be made by Economic Operators.

Therefore, as the above answer to the question shall not cause modification of the Specification the extension of offer submission time limit is not necessary.

Moreover, referring to your demand, even if Economic Operator had such a chart or diagram at his disposal he would be obliged to put these on his website pursuant to the principles specified in Article 7 of Public Procurement Law of equal treatment of all the Economic Operators, which would convey in such a case an equal access to information for all the Economic Operators.

z up. DYREKTOR

mgr Jolanta Kobus
Przewodnicząca Komisji Przetargowej